


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>4110</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Cambridge Vibration Maintenance Services Ltd</h3> <p>Issue No: 003 Issue date: 13 March 2012</p>	
	<p>Millside The Moor Melbourn Herts SG8 6ED</p>	<p>Contact: Graham Hagger Tel: +44 (0)1763 262 112 Fax: +44 (0)1763 263 335 E-Mail: sales@cvmsl.co.uk Website: www.cvmsl.co.uk</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
<p>ACCELERATION TRANSDUCERS</p> <p>Reference (precision) Grade</p> <p>Piezoelectric type Transducer at 23°C</p>	<p>High frequency Test Nominal peak acceleration $1g_n$ up to $10g_n$ (9.81 up to 98 m/s²) Charge Sensitivity >0.1 pCg_n (0.01 pC/ms⁻²) <1000 pCg_n (0.01 pC/ms⁻²)</p> <p>20 Hz to 5 kHz 5 kHz to 6.3 kHz 6.3 kHz to 10 kHz</p>	<p>1.5 % 2.0 % 2.5 %</p>	<p>Calibration of charge sensitivity by comparison with a reference (precision grade) transducer</p> <p>Transducer at ambient</p>
<p>Working or non-precision grades Piezoelectric type</p>	<p>High frequency Test Nominal peak acceleration $1g_n$ up to $10g_n$ (9.81 up to 98 m/s²) Charge Sensitivity >0.1 pCg_n (0.01 pC/ms⁻²) <1000 pCg_n (0.01 pC/ms⁻²)</p> <p>20 Hz to 5 kHz 5 kHz to 6.3 kHz 6.3 kHz to 10 kHz</p> <p>Low frequency Test Nominal peak acceleration 0.2 up to 2 g_n (1.96 up to 19.6 m/s²) Charge Sensitivity >2 pCg_n (0.01 pC/ms⁻²) <1000 pCg_n (0.01 pC/ms⁻²)</p> <p>2 Hz to 20 Hz</p>	<p>1.5 % 2.0 % 2.5 % 1.5 %</p>	<p>Calibration of charge sensitivity by comparison with a reference (precision grade) transducer</p> <p>Transducer at ambient</p>



4110
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Cambridge Vibration Maintenance Services Ltd
Issue No: 003 Issue date: 13 March 2012

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k=2)	Remarks
ACCELERATION TRANSDUCERS			Calibration of sensitivity by comparison with a reference (precision grade) transducer
Piezoresistive or strain-gauge type	High frequency Test Nominal peak acceleration 1 up to 10 g _n (9.81 up to 98 m/s ⁻²) System Sensitivity > 1mv/g _n (0.1 mv/ms ⁻²) < 1000 mv/g _n (100 mv/ms ⁻²)		Transducer at ambient
	20 Hz to 5 kHz 5 kHz to 6.3 kHz 6.3 kHz to 10 kHz	1.5 % 2.0 % 2.5 %	
	Low frequency Test Nominal peak acceleration 0.2 up to 2 g _n (1.96 up to 19.6 m/s ⁻²) System Sensitivity >0.05 mv/g _n (0.005 mv/ms ⁻²) < 1000 mv/g _n (100 mv/ms ⁻²)		Transducer at ambient
Integral electronics type	2 Hz to 20 Hz	1.5 %	Calibration of sensitivity by comparison with a reference (precision grade) transducer
	High frequency Test Nominal peak acceleration 1 up to 10 g _n (9.81 up to 98 m/s ⁻²) System Sensitivity > 1 mv/g _n (0.1 mv/ms ⁻²) < 1000 mv/g _n (100 mv/ms ⁻²)		Transducer at ambient
	20 Hz to 5 kHz 5 kHz to 6.3 kHz 6.3 kHz to 10 kHz	1.5 % 2.0 % 2.5 %	
	Low frequency Test Nominal peak acceleration 0.2 up to 2 g _n (1.96 up to 19.6 m/s ⁻²) System Sensitivity >1mv/g _n (0.1 mv/ms ⁻²) <1000 mv/g _n (100 mv/ms ⁻²)		Transducer at ambient
	2 Hz to 20 Hz	1.5 %	
PORTABLE ACCELEROMETER CALIBRATORS	Over the ranges detailed above.		



4110
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Cambridge Vibration Maintenance Services Ltd
Issue No: 003 Issue date: 13 March 2012

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
SUPPORTING MEASUREMENTS			
AC CURRENT Generation	2.0 mA to 20 mA 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz	0.5 % + 4.0 μ A 0.3 % + 4.0 μ A 0.4 % + 4.0 μ A	Ancillary measurements associated with Accelerometry calibration only
TEMPERATURE	18 °C to 28 °C	1.0 °C	
END			